

Technical Specification Of CR System

Purchaser : Aathbiskot Municipality, Rukum(West)

Medical Equipment : Digital X-ray

Quantity : 1 Unit




S.N.	CRITERIA	OFFICE REQUIREMENT	OFFERED SPECIFICATION
1	CR Reader Unit	<ul style="list-style-type: none"> ❖ Reading Capacity : Minimum 55 plates/hr. ❖ Vertical Cassette Insertion ❖ OPG(Ortho Pantomography) Readable ❖ Operatable on Single Phase Line ❖ Non Table-Top Model 	
2	CR Console	<ul style="list-style-type: none"> • Main Workstation-Multi Function CR Console with assured quality • High Resolution 19" LED Monitor • Adequate Image Capacity of 500 GB • Multi Patient Viewing & Printing • Image Magnification • Distance & Angle Measurement • Exporting images in BMP, JPEG and AVI Formats on DICOM & NON-DICOM • Burning Facility in CD & DVD 	
3	CR Cassettes & Images Plates	<ul style="list-style-type: none"> ◆ 8"×10"=1 ◆ 10"×12"=1 ◆ 14"×14"=1 	
4	Dry Laser Printer	<ul style="list-style-type: none"> ✓ Dry Laser Printer ✓ Printing Capacity : Minimum 80 sheets/hr. of 14"×17" ✓ Pixel sizes : Acquiring both 50 μm(508 dpi) & 100 μm(254 dpi) ✓ N0. of Trays M 2 trays online ✓ Compatible Film Type : Dry Film Sensitive to Light ✓ Film Print Size : 8"×10", 10"×12", 10"×14", 14"×14", 14"×17" ✓ Daylight Film Loading System ✓ Minimum Grayscale Resolution of 14 Bits ✓ Operatable on Single Phase Line ✓ Non Table-Top Model 	

Note:

The major items are indicated by bold and highlighted letters/alphabets/figures in the technical specifications. Non-compliance of these major items will result in non-responsiveness of the bid.

Bidder's Signature.....

Date.....

A handwritten signature in black ink, appearing to be a stylized 'W' or 'V' followed by a large loop and a horizontal stroke.